

Certificate of Quality

xMAP® INTELLIFLEX Performance Verification Kit, RUO IFX-PVER-K20

REF Product Number IFX-PVER-K20
LOT Lot Number C18113
 Kit Expiration Date 2027-04-30

Kit Components	Part number	Lot*
xMAP® INTELLIFLEX Verifier, Microspheres, 10 mL	IFXVER-10	C16784
xMAP® Fluidics 1, Microspheres, 5 mL	FLUID1-05	C15426
xMAP® Fluidics 2, Microspheres, 5 mL	FLUID2-05	C15323
8-Well Microtiter Plate Strip Wells, Pack of 25	13-52047	N/A
INTELLIFLEX Performance Verification Kit Target Values Card	89-30000-00-889	C18113

	DD Target Value	CL1 Target Value	CL2 Target Value	CL3 Target Value	RP1 Target Value	RP2 Target Value
Reference 6	9206	57.5	68.0	24.9	N/A	N/A
Reference 22	9099	1295	231	65.1	7.3	207
Reference 71	9126	57.2	4823	2165	48070	N/A
Reference 81	10674	21951	5567	1758	N/A	24971
Reference 96	8785	2426	13207	6744	N/A	29.3
Reference 102	9097	58.6	88.4	155	16916	N/A
Reference 153	8769	55.0	4900	9672	N/A	1592
Reference 352	10465	690	1297	2511	2227	N/A
Reference 374	10387	666	1616	9056	32.8	N/A
Reference 378	9116	1211	361	359	186	15788
Reference 500	8966	12937	8122	9091	N/A	N/A

This certificate provides descriptive lot information and assurance that quality requirements have been met for the components, intermediates, and finished product. Luminex Corporation certifies that the xMAP® INTELLIFLEX Performance Verification Kit, IFX-PVER-K20, lot defined above conforms to all prescribed acceptance criteria and is fit for its intended use.



Amy Branecky
2025.11.26 10:45:31 -06'00'

Senior Manager, Reagent Quality Control

*Reagent target values may vary lot to lot. Only the reagent lots listed should be used when performing system verification with this kit lot.

Luminex
EN ISO 13485 Certified Company

12212 Technology Blvd • Austin, Texas 78727-6115 USA

Phone: 1.512.381.4397 • Toll-free: 1.877.785.2323 • Fax: 512.219.5195 • www.luminexcorp.com